


ASSAY VALUES AND EXPECTED RANGES

QCP DATA MONTHS: FEBRUARY, MARCH

ADVIA® 120, 2120 & 2120i LEVEL LOW

Parameter	Mean	Range
WBC × 10 ³ /μL	3.20	± 0.50
WBCP PEROX × 10 ³ /μL	2.75	± 0.50
RBC × 10 ⁶ /μL	2.04	± 0.12
HGB g/dL	5.5	± 0.4
HCT %	15.0	± 1.5
MCV fL	73.5	± 4.0
MCH pg	27.0	± 1.8
MCHC g/dL	36.7	± 2.5
CHCM g/dL	35.5	± 3.0
RDW %	16.0	± 3.0
HDW g/dL	1.90	± 1.00




LOT FE6614



2024-04-05

Parameter	Mean	Range
PLT × 10 ³ /μL	58	± 20
MPV fL	9.0	± 3.0
NEUT %	43.0	± 13.0
LYMP %	33.0	± 14.0
MONO %	14.5	± 10.0
EOS %	2.0	± 2.0
BASO %	1.5	± 1.5
LUC %	6.0	± 6.0
NEUTx	NA	NA
NEUTy	NA	NA



NOTE:

An occasional test may exhibit Lymp% and Neut% imprecision. If this occurs, repeat test. If it persists, call Hematology Technical Service at 1-800-523-3395.

Before using, refer to the instruction sheet for mixing directions.
For technical assistance in the USA and Canada call Technical Service at 1-800-523-3395.
CBC-Tech must be analyzed in the instrument Quality Control Mode.
All brands and products are trademarks or registered trademarks of their respective companies.


ASSAY VALUES AND EXPECTED RANGES

QCP DATA MONTHS: FEBRUARY, MARCH

++

ADVIA 120, 2120 & 2120i LEVEL NORMAL

Parameter	Mean	Range
WBC × 10 ³ /μL	7.60	± 1.00
WBCP PEROX × 10 ³ /μL	7.20	± 1.00
RBC × 10 ⁶ /μL	4.15	± 0.20
HGB g/dL	12.6	± 0.5
HCT %	34.4	± 2.6
MCV fL	83.0	± 4.0
MCH pg	30.4	± 1.9
MCHC g/dL	36.6	± 2.8
CHCM g/dL	35.0	± 3.0
RDW %	14.5	± 3.0
HDW g/dL	1.55	± 1.00




LOT FE6624



2024-04-05

Parameter	Mean	Range
PLT × 10 ³ /μL	221	± 45
MPV fL	8.9	± 3.0
NEUT %	63.0	± 13.0
LYMP %	20.5	± 12.0
MONO %	7.5	± 7.5
EOS %	3.0	± 3.0
BASO %	1.0	± 1.0
LUC %	5.0	± 5.0
NEUTx	60.0	± 6.0
NEUTy	80.0	± 8.0



NOTE: An occasional test may exhibit Lymph% and Neut% imprecision. If this occurs, repeat test. If it persists, call Hematology Technical Service at 1-800-523-3395.

Before using, refer to the instruction sheet for mixing directions.
For technical assistance in the USA and Canada call Technical Service at 1-800-523-3395.
CBC-Tech must be analyzed in the instrument Quality Control Mode.
All brands and products are trademarks or registered trademarks of their respective companies.

ASSAY VALUES AND EXPECTED RANGES

QCP DATA MONTHS: FEBRUARY, MARCH

+++

ADVIA 120, 2120 & 2120i LEVEL HIGH

LOT FE6634



2024-04-05

Parameter	Mean	Range
WBC × 10 ³ /μL	17.20	± 2.80
WBCP PEROX × 10 ³ /μL	17.20	± 2.80
RBC × 10 ⁶ /μL	5.08	± 0.24
HGB g/dL	16.5	± 0.7
HCT %	45.2	± 3.5
MCV fL	89.0	± 4.0
MCH pg	32.5	± 2.0
MCHC g/dL	36.5	± 3.0
CHCM g/dL	34.0	± 3.0
RDW %	13.5	± 3.0
HDW g/dL	1.45	± 1.00

Parameter	Mean	Range
PLT × 10 ³ /μL	430	± 65
MPV fL	8.9	± 3.0
NEUT %	70.0	± 12.0
LYMP %	15.0	± 10.0
MONO %	7.5	± 7.5
EOS %	3.5	± 3.5
BASO %	1.0	± 1.0
LUC %	3.0	± 3.0
NEUTx	59.0	± 6.0
NEUTy	80.5	± 8.0

NOTE: An occasional test may exhibit Lymp% and Neut% imprecision. If this occurs, repeat test. If it persists, call Hematology Technical Service at 1-800-523-3395.

Before using, refer to the instruction sheet for mixing directions.
For technical assistance in the USA and Canada call Technical Service at 1-800-523-3395.
CBC-Tech must be analyzed in the instrument Quality Control Mode.
All brands and products are trademarks or registered trademarks of their respective companies.